

**DR. S. S. BHATNAGAR UNIVERSITY INSTITUTE OF CHEMICAL
ENGINEERING AND TECHNOLOGY, PANJAB UNIVERSITY, CHANDIGARH**

No. : 3363/UICET

Date: 12.01.2023

Sub: Inviting Quotations for Specifications for D2 Lamp for UV-VIS Spectrophotometer

Dear Sir/Madam,

You are invited to submit your most competitive quotation for Specifications for D2 Lamp for UV-VIS Spectrophotometer.

Specification	
Electrode Structure	Standard
Window Material	UV Glass
Spectral Distribution (Short)	Up to 185 nm
Spectral Distribution (Long)	Up to 400 nm
Aperture Diameter *a	0.5 mm
Output Stability Drift Max.	+/- 0.3 %/h
Output Stability Fluctuation (p-p) Typ.	0.005 %
Guaranteed Life *d	1500-2000 h
Filament Warm-up Time	15-20 s
Filament Operating Voltage	1.0 +/- 0.1 V dc

Terms and Conditions:

- 1) Sealed quotation should reach the on the below address latest by **27.01.2023** by **12.00 noon**. only: Chairperson, Dr. S. S. Bhatnagar University Institute of Chemical Engineering & Technology, Panjab University, Chandigarh-160014
- 2) Envelope containing the quotation should be super scribed "Quotation for Specifications for D2 Lamp for UV-VIS Spectrophotometer".
- 3) Rates should be quoted inclusive of all the taxes.
- 4) No advancement payment will be made.
- 5) The firm should quote lowest rate of described work individually and total value.
- 6) Conditional quotation will not be accepted.

- 7) The undersigned has the right to accept or reject any or all the quotations without assigning any reason.
- 8) Tenders received after the due date will not be accepted. If the last date for submission of tender falls on any declared holiday in the University, the next working day will be considered as the last date for the same.
- 9) Telegraphic/electronic/conditional offers will not be accepted.
- 10) **Site visit should be done for performing Installation and calibration of D2 Lamp.**

Chairperson