

Addendum in the specifications of Balances 0.1mg:

PLEASE REFER THIS BELOW REVISED SPECIFICATIONS FOR QUOTING THE MODEL OF BALANCES:

Capacity : **220g or more**

Minimum Weighing Value : 0.1 mg or less

Repeatability : 0 to 200g; 0.1mg

200 to 252g; 0.2 mg

Linearity : ± 0.3 mg

Stabilization Time: app. 2 sec

Clear, Reverse Backlit LCD

One-Touch Automatic Calibration

Sensitivity Drift: ± 2 ppm/ $^{\circ}$ C (10 to 30 $^{\circ}$ C/50 to 86 $^{\circ}$ F)

GLP/GMP/GCP/ISO Compliant Output

Rotary Sliding Doors

Weighing Pan : dia 90mm

Internal Calibration

Clock & calendar function

Coordinator TEQIP-II